



Biochrom Ltd  
Certificate No. 890333

## CERTIFICATE OF ANALYSIS

for

### Ultra Ninhydrin Solution

Part Number: 80-2117-64

Control Number: 19076

Quantity: 250mL

The above product gave the following results on analysis:

Base line drift no more than:

-18.7 % at 440nm on 0.1 OD range

-2.5 % at 570nm on 0.1 OD range

Chemical QC

Date: 03 February 2012

---

|  |   |                    |                        |
|--|---|--------------------|------------------------|
| Postal Address   | Telephone   | Telefax            | Email:                 |
| Biochrom Ltd<br>22 Cambridge Science Park<br>Milton Road<br>Cambridge CB4 0FJ<br>England | +44 (0)1223 423723  | +44 (0)1223 420164 | support@biochrom.co.uk |
|  | Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a> |                    |                        |

Registered in England No: 3526954 Registered Office: 22 Cambridge Science Park, Milton Road, Cambridge, CB4 0FJ, England.