

Declaration of CE Conformity

Manufacturer: Biochrom Ltd
Address: 22 The Science Park, Cambs, CB4 0FJ, UK
Type of Equipment: Laboratory Equipment
Product description: **Microplate Reader MultiRead 400, Order number G033100**

The described product meets the essential requirements set out in the following Directives

73/23/EEC & 89/336/EEC

and Directive 98 / 79 / EC of the European Parliament and the Council of 27 Oct. 1998 on In Vitro Diagnostic Medical Devices.

The described product meets the essential requirements set out in Annex I of Directive 98/79/EC which apply to it, according to the intended purpose of the product.

The manufacturer declares that the technical documentation of the product fulfils the requirements set out in Annex III, section 3 of Directive 98/79/EC.

Standards against which conformity is assessed:

EN 61010-1: (1993)
EN 61326 (1998)
IEC61000-4-2: (2001)
IEC61000-3-3: (1994)
IEC61000-4-4: (2004)
IEC61000-4-11: (2004)
IEC61000-4-5: (2006)
EN 61010-2-101: (2002)
EN 1658; 1997

Original Issuer: Biochrom Ltd
Place, Date: Cambridge, July 1st, 2008

A handwritten signature in black ink, appearing to be 'S P Marriott', written over a faint horizontal line.

Signature:

S P Marriott
R&D Director, Biochrom

This declaration confirms the agreement of the mentioned guidelines, comprises no assurance of features.
Please follow the safety notes of the included product information.