

# CERTIFICATE OF ANALYSIS

For

## Ultra Ninhydrin Solution NC

Part Number: 80-6003-08

Batch (Control) Number: P27738NC

Quantity: 250 mL

Expiry Date: 18 March 2024

The above product gave the following results on analysis:

### Baseline Drift:

0.019  $\Delta mV/\Delta t$  at 440nm, 0.026  $\Delta mV/\Delta t$  at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

### Colorimetric Test:

Absorption @ 485nm = 0.034 A

Absorption @ 600nm = 0.025 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Date of Issue: 18 March 2021

David Thomas

---

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.