

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 2

Part Number: 80-2115-22

Batch (Control) Number: P27762

Quantity: 2 Litres

Expiry Date: 04 August 2024

The above product gave the following results on analysis:

pH = 3.504 @ 20.0°C

pH Specification = 3.48 - 3.52 @ 20.0°C

Conductivity = 10.36 mScm⁻¹ @ 20.0°C

Conductivity Specification = 9.5 - 11.5 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 10 August 2021

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.