

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 4

Part Number: 80-2097-18

Batch (Control) Number: R27785

Quantity: 2 Litres

Expiry Date: 10 January 2025

The above product gave the following results on analysis:

pH = 3.494 @ 20.0°C pH Specification = 3.48 - 3.52 @ 20.0°C

Conductivity = 57.6 mScm-1 @ 20.0°C Conductivity Specification = 54.9 - 58.2 mScm-1 @ 20.0°C

Chemical QC Signature: Neil Booth

Waterbeach, Cambridge, CB25 9PE England.

Date of Issue: 28 January 2022

 Postal Address
 Telephone
 Telefax
 Email:

 Biochrom Ltd
 +44 (0)1223 423723
 +44 (0)1223 420164
 support@biochrom.co.uk

 Unit 7, Enterprise Zone
 3970 Cambridge Research Park
 support@biochrom.co.uk

 Beach Drive, Waterbeach
 Cambridge
 CB25 9PE, England.
 Web Site: http://www.biochrom.co.uk

 Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive,
 Support @biochrom.co.uk

Certificate of Analysis v1.1 - 29 September 2020.