

CERTIFICATE OF ANALYSIS

For

EZ Nin Reagent NC

Part Number: 80-6003-19

Batch (Control) Number: R27802NC

Quantity: 2 Litres

Expiry Date: 5 May 2025

The above product gave the following results on a standard Lithium High-Performance System:

Baseline Drift:

-0.030 $\Delta mV/\Delta t$ at 440nm, 0.002 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Chemical QC Signature:
David Thomas

Date of Issue: 10 May 2022

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
	Web Site: http://www.biochrom.co.uk		

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.