

CERTIFICATE OF ANALYSIS

For

Sodium Hydrolysate Buffer 1

Part Number: 80-2037-68

Batch (Control) Number: T27835

Quantity: 2 Litres

Expiry Date: 09 February 2026

The above product gave the following results on analysis:

pH = 3.200 @ 20.0°C pH Specification = 3.18 - 3.22 @ 20.0°C

Conductivity = 15.10 mScm-1 @ 20.0°C Conductivity Specification = 14.4 - 16.1 mScm-1 @ 20.0°C

Chemical QC Signature: Neil Booth

Date of Issue: 10 February 2023

 Postal Address
 Telephone
 Telefax
 Email:

 Biochrom Ltd
 +44 (0)1223 423723
 +44 (0)1223 420164
 support@biochrom.co.uk

 Unit 7, Enterprise Zone
 3970 Cambridge Research Park
 support@biochrom.co.uk

 Beach Drive, Waterbeach
 Cambridge
 CB25 9PE, England.
 Web Site: http://www.biochrom.co.uk

 Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive,
 Postal Address

Waterbeach, Cambridge, CB25 9PE England. Certificate of Analysis v1.1 - 29 September 2020.