

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 5 NC

Part Number: 80-6003-05

Batch (Control) Number: T27857NC

Quantity: 2 Litres

Expiry Date: 29 June 2026

The above product gave the following results on analysis:

pH = 3.543 @ 20.0°C pH Specification = 3.53 - 3.57 @ 20.0°C

Conductivity = 88.8 mScm-1 @ 20.0°C Conductivity Specification = 82.8 - 89.6 mScm-1 @ 20.0°C

Chemical QC Signature: Date of Issue: 17 July 2023

David Thomas

Postal Address Telephone

+44 (0)1223 423723

Telefax +44 (0)1223 420164 Email:

support@biochrom.co.uk

Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach

Cambridge

Biochrom Ltd

CB25 9PE, England. Web Site: http://www.biochrom.co.uk

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.

Certificate of Analysis v1.1 - 29 September 2020.