

# CERTIFICATE OF ANALYSIS

For

## Sodium Oxidized Buffer 4

Part Number: 80-2037-88

Batch (Control) Number: R27829

Quantity: 2 Litres

Expiry Date: 19 December 2025

The above product gave the following results on analysis:

pH = 8.592 @ 20.0°C

pH Specification = 8.58 - 8.62 @ 20.0°C

Conductivity = 38.2 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 35.8 - 41.4 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:

Date of Issue: 21 December 2022

Neil Booth

---

Postal Address

Telephone

Telefax

Email:

Biochrom Ltd  
Unit 7, Enterprise Zone  
3970 Cambridge Research Park  
Beach Drive, Waterbeach  
Cambridge  
CB25 9PE, England.

+44 (0)1223 423723

+44 (0)1223 420164

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.